

Alpha Resources, Inc. Certificate Of Analysis

AR2712

CALCINATED PETROLEUM COKE STANDARD REFERENCE MATERIAL

LOT #7120497

WEIGHT PERCENT SULFUR = 0.43
STANDARD DEVIATION = 0.04

The certified mean values were statistically derived from the values reported to Alpha Resources through Round Robin.

Methods

ASTM D 4239 METHOD B
ASTM D 4239 METHOD C

Notes:

Normal ASTM procedures should be employed when using this standard. This includes using the *reproducibility* and *repeatability* factors. Use the reproducibility factor when establishing the first set of data, and then establish an internal repeatability factor as your guideline.

The statistical analysis of the certification was performed by D.M. Reasons at Alpha Resources.

The overall direction and coordination of the analytical measurements leading to certification were performed by D.M. Reasons at Alpha Resources.

Remedies for any claimed defect in this product will be limited to product replacement or refund of the purchase price. In no event shall Alpha Resources be liable for incidental or consequential damages.

EXPIRATION DATE

THIS CRM IS VALID FOR TWO YEARS FROM THE DATE OF OPENING



P.O Box 199 3090 Johnson Road Stevensville, MI 49127 Phone (269) 465-5559 Fax (269) 465-3629

www.alpharesources.com

Page 1 Of 1